


CAM-SERIES

compatible to:

Altatec[®]

Camlog[®]



A: over surface

B: over edge

PLATFORMS:





TITANIUM BASE

incl. screw

CAM 80330000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

CAM 9.S3D3.300



NT-MILL LAB ANALOG

CAM 5.La3.300



IMPLANT PICK-UP

for open-tray with integrated screw

CAM TR-00013.3



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

CAM 6033843



DIM ANALOG®

CAM 5.DIM.330



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM 5.XSL.A10





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>CAM 10.Mi3.300.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>CAM 9.S3D3.300</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>CAM 5.La3.300</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>CAM 6.Sc3.384</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>CAM 6033843</p>	
<p>DIM ANALOG®</p>	<p>CAM 5.DIM.330</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A10</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.100.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

CAM
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	CAM-2CON 8033001S
GH 2,5 mm	CAM-2CON 8033002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	CAM-2CON 8033001M
GH 2,5 mm	CAM-2CON 8033002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm


A: over surface CAM-2CON 80330C3S
B: over edge CAM-2CON 80330A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

A: over surface CAM-2CON 80330C3M
B: over edge CAM-2CON 80330A3M



CAM
SERIES

2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

CAM 6.Ms3.384



2-CONNECT®-DIM ANALOG®

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600



CAM-SERIES

3,8 mm



TITANIUM BASE

incl. screw

CAM 80380000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

CAM 9.S3D3.800



NT-MILL LAB ANALOG

CAM 5.La3.800



IMPLANT PICK-UP

for open-tray with integrated screw

CAM TR-00023.8



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

CAM 6033843



DIM ANALOG®

CAM 5.DIM.380



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A40



CAM
SERIES



<p>NT-PREFORM® incl. coloured milling screw</p>	<p>CAM 10.Mi3.800.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>CAM 9.S3D3.800</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>CAM 5.La3.800</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>CAM 6.Sc3.384</p>	
<p>ABUTMENT SCREW  Overview of all screwdrivers on pages 344 and 345</p>	<p>CAM 6033843</p>	
<p>DIM ANALOG®</p>	<p>CAM 5.DIM.380</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A40</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.100.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p style="color: red; font-size: small;">Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

CAM
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm CAM-2CON 8038001S

GH 2,5 mm CAM-2CON 8038002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm CAM-2CON 8038001M

GH 2,5 mm CAM-2CON 8038002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800






2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

A: over surface CAM-2CON 80380C3S
B: over edge CAM-2CON 80380A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

A: over surface CAM-2CON 80380C3M
B: over edge CAM-2CON 80380A3M



CAM
SERIES

2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

CAM 6.Ms3.384



2-CONNECT®-DIM ANALOG

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600





TITANIUM BASE

incl. screw

CAM 80430000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

CAM 9.S3D4.300



NT-MILL LAB ANALOG

CAM 5.La4.300



IMPLANT PICK-UP

for open-tray with integrated screw

CAM TR-00034.3



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

CAM 6033843



DIM ANALOG®

CAM 5.DIM.430



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A40





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>CAM 10.Mi4.300.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>CAM 9.S3D4.300</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>CAM 5.La4.300</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>CAM 6.Sc3.384</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>CAM 6033843</p>	
<p>DIM ANALOG®</p>	<p>CAM 5.DIM.430</p>	
<p>SLEEVE (belongs to DIM ANALOG) plaster model</p>	<p>DIM 5.XSL.A40</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

CAM
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	CAM-2CON 8043001S
GH 2,5 mm	CAM-2CON 8043002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	CAM-2CON 8043001M
GH 2,5 mm	CAM-2CON 8043002M



CAM
SERIES

2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

A: over surface CAM-2CON 80430C3S
B: over edge CAM-2CON 80430A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

A: over surface CAM-2CON 80430C3M
B: over edge CAM-2CON 80430A3M



CAM
SERIES

2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

CAM 6.Ms3.384



2-CONNECT®-DIM ANALOG

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600



CAM-SERIES

5,0 mm



TITANIUM BASE

incl. screw

CAM 80500000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

CAM 9.S3D5.060



NT-MILL LAB ANALOG

CAM 5.La5.060



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

CAM 6150600



DIM ANALOG®

CAM 5.DIM.506



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A50



CAM
SERIES



<p>NT-PREFORM® incl. coloured milling screw</p>	<p>CAM 10.Mi5.000.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>CAM 9.S3D5.060</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>CAM 5.La5.060</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>CAM 6.Sc5.060</p>	
<p>ABUTMENT SCREW</p> <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>CAM 6150600</p>	
<p>DIM ANALOG®</p>	<p>CAM 5.DIM.506</p>	
<p>SLEEVE (belongs to DIM ANALOG) plaster model</p>	<p>DIM 5.XSL.A50</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.300.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

CAM
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	CAM-2CON 8050001S
GH 2,5 mm	CAM-2CON 8050002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	CAM-2CON 8050001M
GH 2,5 mm	CAM-2CON 8050002M



CAM
SERIES

2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)	A: over surface B: over edge	CAM-2CON 80500C3S CAM-2CON 80500A3S	
2-CONNECT®-ABUTMENT angled 20°, Gingiva Height 2,5 mm Male Part incl. screw, mounting aid	A: over surface B: over edge	CAM-2CON 80500C3M CAM-2CON 80500A3M	
2-CONNECT®-SCAN BODY 3D-GUIDE® for intra + extra oral scanning KS 62 = 1,6 mm		2CON 9.S3D1.600	
2-CONNECT®-COPING Female Part		2CON F-U	
COPING-SCREW Recommended tightening torques: 15 Ncm		KS 62	
MALE PART SCREW for angled 20° 2-CONNECT® Abutment Recommended tightening torques: 20 Ncm		CAM 6.Ms5.060	
2-CONNECT®-DIM ANALOG® KS 62 = 1,6 mm		2CON 5.DIM.160	
SLEEVE (belongs to DIM ANALOG®) plaster model		DIM 5.XSL.A30	
2-CONNECT®-IMPLANT PICK-UP integrated screw KS 62 = 1,6 mm		2CON TR-00031.6	
2-CONNECT®-LAB ANALOG KS 62 = 1,6 mm		2CON 5.La1.600	