


BIO-SERIES

compatible to:

BioHorizons[®] / internal



PLATFORMS:



3,0 mm



3,5 mm



4,5 mm



5,7 mm



TITANIUM BASE

incl. screw

BIO 80300000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BIO 9.S3D3.000



NT-MILL LAB ANALOG

BIO 5.La3.000



IMPLANT PICK-UP

for open-tray with integrated screw

BIO TR-00013.0



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BIO 603057



DIM ANALOG®

BIO 5.DIM.300



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM XSL.A10





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BIO 10.Mi3.000.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BIO 9.S3D3.000</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BIO 5.La3.000</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BIO 6.Sc3.057</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BIO 603057</p>	
<p>DIM ANALOG®</p>	<p>BIO 5.DIM.300</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A10</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.100.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm BIO-2CON 8030001S

GH 2,5 mm BIO-2CON 8030002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm BIO-2CON 8030001M

GH 2,5 mm BIO-2CON 8030002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

BIO-2CON 80300A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

BIO-2CON 80300A3M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

BIO 6.Ms3.057



2-CONNECT®-DIM ANALOG®

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600





TITANIUM BASE

incl. screw

BIO 80350000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BIO 9.S3D3.500



NT-MILL LAB ANALOG

BIO 5.La3.500



IMPLANT PICK-UP

for open-tray with integrated screw

BIO TR-00023.5



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BIO 603057



DIM ANALOG®

BIO 5.DIM.350



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM 5.XSL.A10





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BIO 10.Mi3.500.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BIO 9.S3D3.500</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BIO 5.La3.500</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BIO 6.Sc3.057</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BIO 603057</p>	
<p>DIM ANALOG®</p>	<p>BIO 5.DIM.350</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A10</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.100.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm BIO-2CON 8035001S

GH 2,5 mm BIO-2CON 8035002S



BIO
SERIES

2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm BIO-2CON 8035001M

GH 2,5 mm BIO-2CON 8035002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





<p>2-CONNECT®-ABUTMENT SET (ANGLED) incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	BIO-2CON 80350A3S	
<p>2-CONNECT®-ABUTMENT angled 20°, Gingiva Height 2,5 mm Male Part incl. screw, mounting aid</p>	BIO-2CON 80350A3M	
<p>2-CONNECT®-SCAN BODY 3D-GUIDE® for intra + extra oral scanning KS 62 = 1,6 mm</p>	2CON 9.S3D1.600	
<p>2-CONNECT®-COPING Female Part</p>	2CON F-U	
<p>COPING-SCREW Recommended tightening torques: 15 Ncm</p>	KS 62	
<p>MALE PART SCREW for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	BIO 6.Ms3.057	
<p>2-CONNECT®-DIM ANALOG® KS 62 = 1,6 mm</p>	2CON 5.DIM.160	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	DIM 5.XSL.A30	
<p>2-CONNECT®-IMPLANT PICK-UP integrated screw KS 62 = 1,6 mm</p>	2CON TR-00031.6	
<p>2-CONNECT®-LAB ANALOG KS 62 = 1,6 mm</p>	2CON 5.La1.600	



TITANIUM BASE

incl. screw

BIO 80450000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BIO 9.S3D4.500



NT-MILL LAB ANALOG

BIO 5.La4.500



IMPLANT PICK-UP

for open-tray with integrated screw

BIO TR-00034.5



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BIO 603057



DIM ANALOG®

BIO 5.DIM.450



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BIO 10.Mi4.500.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BIO 9.S3D4.500</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BIO 5.La4.500</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BIO 6.Sc3.057</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BIO 603057</p>	
<p>DIM ANALOG®</p>	<p>BIO 5.DIM.450</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	BIO-2CON 8045001S
GH 2,5 mm	BIO-2CON 8045002S



BIO
SERIES

2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	BIO-2CON 8045001M
GH 2,5 mm	BIO-2CON 8045002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

BIO-2CON 80450A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

BIO-2CON 80450A3M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

BIO 6.Ms3.057



2-CONNECT®-DIM ANALOG®

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600





TITANIUM BASE

incl. screw

BIO 80570000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BIO 9.S3D5.700



NT-MILL LAB ANALOG

BIO 5.La5.700



IMPLANT PICK-UP

for open-tray with integrated screw

BIO TR-00045.7



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BIO 603057



DIM ANALOG®

BIO 5.DIM.570



SLEEVE (belongs to DIM ANALOG®)
plaster model

DIM 5.XSL.A60





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BIO 10.Mi5.700.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BIO 9.S3D5.700</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BIO 5.La5.700</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BIO 6.Sc3.057</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BIO 603057</p>	
<p>DIM ANALOG®</p>	<p>BIO 5.DIM.570</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A60</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.300.000</p>	
<p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm BIO-2CON 8057001S

GH 2,5 mm BIO-2CON 8057002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm BIO-2CON 8057001M

GH 2,5 mm BIO-2CON 8057002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





2-CONNECT®-ABUTMENT SET (ANGLED)

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

BIO-2CON 80570A3S



2-CONNECT®-ABUTMENT

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

BIO-2CON 80570A3M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 62



MALE PART SCREW

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

BIO 6.Ms3.057



2-CONNECT®-DIM ANALOG®

KS 62 = 1,6 mm

2CON 5.DIM.160



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6



2-CONNECT®-LAB ANALOG

KS 62 = 1,6 mm

2CON 5.La1.600

