


NEO-SERIES

compatible to:

NEOSS Implant System[®]
NEOSS[™]



PLATFORMS:



3,5 – 5,5 mm



TITANIUM BASE

incl. screw

NEO 80355500



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

NEO 9.S3D3.555



NT-MILL LAB ANALOG

NEO 5.La3.555



IMPLANT PICK-UP

for open-tray with integrated screw

NEO TR-00013.5



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

NEO 603555



DIM ANALOG®

NEO 5.DIM.355




SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20





| | | |
|---|--------------------------|---|
| <p>NT-PREFORM® incl. coloured milling screw</p> | <p>NEO 10.Mi3.555.05</p> |  |
| <p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p> | <p>NEO 9.S3D3.555</p> |  |
| <p>NT-MILL LAB ANALOG for NT-Preform®</p> | <p>NEO 5.La3.555</p> |  |
| <p>COLOURED MILLING SCREW for NT-Preform®</p> | <p>NEO 6.Sc3.555</p> |  |
| <p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p> | <p>NEO 603555</p> |  |
| <p>DIM ANALOG®</p> | <p>NEO 5.DIM.355</p> |  |
| <p>SLEEVE (belongs to DIM ANALOG®) plaster model</p> | <p>DIM 5.XSL.A20</p> |  |
| <p>ABUTMENT GRIP for NT-Holder fixture</p> | <p>W 12.200.000</p> |  |
| <p>CUSTOMIZED ABUTMENT milled by NT-iDent</p> <p>Not available in the USA</p> | <p>TZ15.TiO.Si1</p> |  |

NEO SERIES

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm NEO-2CON 8035551S

GH 2,5 mm NEO-2CON 8035552S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm NEO-2CON 8035551M

GH 2,5 mm NEO-2CON 8035552M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





| | | |
|---|--------------------------|--|
| <p>2-CONNECT®-ABUTMENT SET (ANGLED) incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p> | <p>NEO-2CON 80355A3S</p> | |
| <p>2-CONNECT®-ABUTMENT angled 20°, Gingiva Height 2,5 mm Male Part incl. screw, mounting aid</p> | <p>NEO-2CON 80355A3M</p> | |
| <p>2-CONNECT®-SCAN BODY 3D-GUIDE® for intra + extra oral scanning KS 62 = 1,6 mm</p> | <p>2CON 9.S3D1.600</p> | |
| <p>2-CONNECT®-COPING Female Part</p> | <p>2CON F-U</p> | |
| <p>COPING-SCREW Recommended tightening torques: 15 Ncm</p> | <p>KS 62</p> | |
| <p>MALE PART SCREW for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p> | <p>NEO 6.Ms3.555</p> | |
| <p>2-CONNECT®-DIM ANALOG® KS 62 = 1,6 mm</p> | <p>2CON 5.DIM.160</p> | |
| <p>SLEEVE (belongs to DIM ANALOG®) plaster model</p> | <p>DIM 5.XSL.A30</p> | |
| <p>2-CONNECT®-IMPLANT PICK-UP integrated screw KS 62 = 1,6 mm</p> | <p>2CON TR-00031.6</p> | |
| <p>2-CONNECT®-LAB ANALOG KS 62 = 1,6 mm</p> | <p>2CON 5.La1.600</p> | |