


# UTG-SERIES

compatible to:

*Sweden & Martina  
Implantology™  
Premium TG®*



**PLATFORMS:**



*3,3 - 4,25 mm*



*5,0 mm*



## TITANIUM BASE

incl. screw

UTG 80334200



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

UTG 9.S3D3.342



## NT-MILL LAB ANALOG

UTG 5.La3.342



## IMPLANT PICK-UP

for open-tray with integrated screw

UTG TR-00013.3



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

UTG 603342



## DIM ANALOG®

UTG 5.DIM.334




## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20





<p><b>NT-PREFORM®</b> incl. coloured milling screw</p>	<p>UTG 10.Mi3.342.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>UTG 9.S3D3.342</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>UTG 5.La3.342</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform®</p>	<p>UTG 6.Sc3.342</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>UTG 603342</p>	
<p><b>DIM ANALOG®</b></p>	<p>UTG 5.DIM.334</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A20</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p style="color: orange;">Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



### 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm    UTG-2CON 8033421S

GH 2,5 mm    UTG-2CON 8033422S



### 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm    UTG-2CON 8033421M

GH 2,5 mm    UTG-2CON 8033422M



### 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



### 2-CONNECT®-COPING

Female Part

2CON F-U



### COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



### 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



### SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



### 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



### 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,  
Coping-screw and Male Part Screw, angled 20°,  
Gingiva Height 2,5 mm

UTG-2CON 80334A3S



**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm  
Male Part inclusive Screw,  
mounting aid

UTG-2CON 80334A3M



**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning  
KS 62 = 1,6 mm

2CON 9.S3D1.600



**2-CONNECT®-COPING**

Female Part

2CON F-U



**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62



**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment  
Recommended tightening torques: 20 Ncm

UTG 6.Ms3.342



**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160



**SLEEVE** (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



**2-CONNECT®-IMPLANT PICK-UP**

integrated screw,  
KS 62 = 1,6 mm

2CON TR-00031.6



**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600





## TITANIUM BASE

incl. screw

UTG 80500000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

UTG 9.S3D5.000



## NT-MILL LAB ANALOG

UTG 5.La5.000



## IMPLANT PICK-UP

for open-tray with integrated screw

UTG TR-00025.0



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

UTG 605000



## DIM ANALOG®

UTG 5.DIM.500



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30





<p><b>NT-PREFORM®</b> incl. coloured milling screw</p>	<p>UTG 10.Mi5.000.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>UTG 9.S3D5.000</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>UTG 5.La5.000</p>	 <p>groove</p>
<p><b>COLOURED MILLING SCREW</b> for NT-Preform®</p>	<p>UTG 6.Sc5.000</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>UTG 605000</p>	
<p><b>DIM ANALOG®</b></p>	<p>UTG 5.DIM.500</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.300.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm    UTG-2CON 8050001S

GH 2,5 mm    UTG-2CON 8050002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm    UTG-2CON 8050001M

GH 2,5 mm    UTG-2CON 8050002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800






**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,  
Coping-screw and Male Part Screw, angled 20°,  
Gingiva Height 2,5 mm

UTG-2CON 80500A3S



**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm  
Male Part inclusive Screw,  
mounting aid

UTG-2CON 80500A3M



**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning  
KS 62 = 1,6 mm

2CON 9.S3D1.600



**2-CONNECT®-COPING**

Female Part

2CON F-U



**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62



**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment  
Recommended tightening torques: 20 Ncm

UTG 6.Ms5.000



**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160



**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A30



**2-CONNECT®-IMPLANT PICK-UP**

integrated screw,  
KS 62 = 1,6 mm

2CON TR-00031.6



**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600

