


# THO-SERIES

compatible to:

*Thommen Medical*<sup>®</sup>  
*SPI*<sup>®</sup>



**PLATFORMS:**





## TITANIUM BASE

incl. screw

THO 80350000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

THO 9.S3D3.500



## NT-MILL LAB ANALOG

THO 5.La3.500



## IMPLANT PICK-UP

for open-tray with integrated screw

THO TR-00013.5



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

THO 603500



## DIM ANALOG®

THO 5.DIM.350



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A10



<p><b>NT-PREFORM®</b> incl. coloured milling screw Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 10.Mi3.500.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>THO 9.S3D3.500</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>THO 5.La3.500</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform® Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 6.Sc3.500</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>THO 603500</p>	
<p><b>DIM ANALOG®</b></p>	<p>THO 5.DIM.350</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A10</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.100.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm THO-2CON 8035001S

GH 2,5 mm THO-2CON 8035002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm THO-2CON 8035001M

GH 2,5 mm THO-2CON 8035002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 60 = 1,4 mm

2CON 9.S3D1.400



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 60



## 2-CONNECT®-DIM ANALOG®

KS 60 = 1,4 mm

2CON 5.DIM.140



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 60 = 1,4 mm

2CON 9.S3D1.400



## 2-CONNECT®-LAB ANALOG

KS 60 = 1,4 mm

2CON 5.La1.400





<p><b>2-CONNECT®-ABUTMENT SET (ANGLED)</b> incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	<p>THO-2CON 80350A3S</p>	
<p><b>2-CONNECT®-ABUTMENT</b> angled 20°, Gingiva Height 2,5 mm Male Part inclusive Screw mounting aid</p>	<p>THO-2CON 80350A3M</p>	
<p><b>2-CONNECT®-SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning KS 62 = 1,6 mm</p>	<p>2CON 9.S3D1.600</p>	
<p><b>2-CONNECT®-COPING</b> Female Part</p>	<p>2CON F-U</p>	
<p><b>COPING-SCREW</b> Recommended tightening torques: 15 Ncm</p>	<p>KS 62</p>	
<p><b>MALE PART SCREW</b> for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	<p>THO 6.Ms3.500</p>	
<p><b>2-CONNECT®-DIM ANALOG®</b> KS 62 = 1,6 mm</p>	<p>2CON 5.DIM.160</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>2-CONNECT®-IMPLANT PICK-UP</b> integrated screw, KS 62 = 1,6 mm</p>	<p>2CON TR-00031.6</p>	
<p><b>2-CONNECT®-LAB ANALOG</b> KS 62 = 1,6 mm</p>	<p>2CON 5.La1.600</p>	



## TITANIUM BASE

incl. screw

THO 80400000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

THO 9.S3D4.000



## NT-MILL LAB ANALOG

THO 5.La4.000



## IMPLANT PICK-UP

for open-tray with integrated screw

THO TR-00024.0



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

THO 604060



## DIM ANALOG®

THO 5.DIM.400



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20





<p><b>NT-PREFORM®</b> incl. coloured milling screw Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 10.Mi4.000.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>THO 9.S3D4.000</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>THO 5.La4.000</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform® Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 6.Sc4.060</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>THO 604060</p>	
<p><b>DIM ANALOG®</b></p>	<p>THO 5.DIM.400</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A20</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	THO-2CON 8040001S
GH 2,5 mm	THO-2CON 8040002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	THO-2CON 8040001M
GH 2,5 mm	THO-2CON 8040002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 60 = 1,4 mm

2CON 9.S3D1.400



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 15 Ncm

KS 60



## 2-CONNECT®-DIM ANALOG®

KS 60 = 1,4 mm

2CON 5.DIM.140



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 60 = 1,4 mm

2CON 9.S3D1.400



## 2-CONNECT®-LAB ANALOG

KS 60 = 1,4 mm

2CON 5.La1.400





<p><b>2-CONNECT®-ABUTMENT SET (ANGLED)</b> incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	<p>THO-2CON 80400A3S</p>	
<p><b>2-CONNECT®-ABUTMENT</b> angled 20°, Gingiva Height 2,5 mm Male Part inclusive Screw, mounting aid</p>	<p>THO-2CON 80400A3M</p>	
<p><b>2-CONNECT®-SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning KS 62 = 1,6 mm</p>	<p>2CON 9.S3D1.600</p>	
<p><b>2-CONNECT®-COPING</b> Female Part</p>	<p>2CON F-U</p>	
<p><b>COPING-SCREW</b> Recommended tightening torques: 15 Ncm</p>	<p>KS 62</p>	
<p><b>MALE PART SCREW</b> for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	<p>THO 6.Ms4.060</p>	
<p><b>2-CONNECT®-DIM ANALOG®</b> KS 62 = 1,6 mm</p>	<p>2CON 5.DIM.160</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>2-CONNECT®-IMPLANT PICK-UP</b> integrated screw, KS 62 = 1,6 mm</p>	<p>2CON TR-00031.6</p>	
<p><b>2-CONNECT®-LAB ANALOG</b> KS 62 = 1,6 mm</p>	<p>2CON 5.La1.600</p>	



## TITANIUM BASE

incl. screw

THO 80450000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

THO 9.S3D4.500



## NT-MILL LAB ANALOG

THO 5.La4.500



## IMPLANT PICK-UP

for open-tray with integrated screw

THO TR-00034.5



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

THO 604060



## DIM ANALOG®

THO 5.DIM.450



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20



<p><b>NT-PREFORM®</b> incl. coloured milling screw Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 10.Mi4.500.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>THO 9.S3D4.500</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>THO 5.La4.500</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform® Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 6.Sc4.060</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>THO 604060</p>	
<p><b>DIM ANALOG®</b></p>	<p>THO 5.DIM.450</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A20</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p style="color: orange;">Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	THO-2CON 8045001S
GH 2,5 mm	THO-2CON 8045002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	THO-2CON 8045001M
GH 2,5 mm	THO-2CON 8045002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





<p><b>2-CONNECT®-ABUTMENT SET (ANGLED)</b> incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	<p>THO-2CON 80450A3S</p>	
<p><b>2-CONNECT®-ABUTMENT</b> angled 20°, Gingiva Height 2,5 mm Male Part inclusive Screw, mounting aid</p>	<p>THO-2CON 80450A3M</p>	
<p><b>2-CONNECT®-SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning KS 62 = 1,6 mm</p>	<p>2CON 9.S3D1.600</p>	
<p><b>2-CONNECT®-COPING</b> Female Part</p>	<p>2CON F-U</p>	
<p><b>COPING-SCREW</b> Recommended tightening torques: 15 Ncm</p>	<p>KS 62</p>	
<p><b>MALE PART SCREW</b> for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	<p>THO 6.Ms4.060</p>	
<p><b>2-CONNECT®-DIM ANALOG®</b> KS 62 = 1,6 mm</p>	<p>2CON 5.DIM.160</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>2-CONNECT®-IMPLANT PICK-UP</b> integrated screw, KS 62 = 1,6 mm</p>	<p>2CON TR-00031.6</p>	
<p><b>2-CONNECT®-LAB ANALOG</b> KS 62 = 1,6 mm</p>	<p>2CON 5.La1.600</p>	



## TITANIUM BASE

incl. screw

THO 80500000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

THO 9.S3D5.000



## NT-MILL LAB ANALOG

THO 5.La5.000



## IMPLANT PICK-UP

for open-tray with integrated screw

THO TR-00045.0



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

THO 604060



## DIM ANALOG®

THO 5.DIM.500



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30





<p><b>NT-PREFORM®</b> incl. coloured milling screw Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 10.Mi5.000.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>THO 9.S3D5.000</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>THO 5.La5.000</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform® Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 6.Sc4.060</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>THO 604060</p>	
<p><b>DIM ANALOG®</b></p>	<p>THO 5.DIM.500</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.300.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm THO-2CON 8050001S

GH 2,5 mm THO-2CON 8050002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm THO-2CON 8050001M

GH 2,5 mm THO-2CON 8050002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





<p><b>2-CONNECT®-ABUTMENT SET (ANGLED)</b> incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	<p>THO-2CON 80500A3S</p>	
<p><b>2-CONNECT®-ABUTMENT</b> angled 20°, Gingiva Height 2,5 mm Male Part inclusive Screw, mounting aid</p>	<p>THO-2CON 80500A3M</p>	
<p><b>2-CONNECT®-SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning KS 62 = 1,6 mm</p>	<p>2CON 9.S3D1.600</p>	
<p><b>2-CONNECT®-COPING</b> Female Part</p>	<p>2CON F-U</p>	
<p><b>COPING-SCREW</b> Recommended tightening torques: 15 Ncm</p>	<p>KS 62</p>	
<p><b>MALE PART SCREW</b> for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	<p>THO 6.Ms4.060</p>	
<p><b>2-CONNECT®-DIM ANALOG®</b> KS 62 = 1,6 mm</p>	<p>2CON 5.DIM.160</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>2-CONNECT®-IMPLANT PICK-UP</b> integrated screw, KS 62 = 1,6 mm</p>	<p>2CON TR-00031.6</p>	
<p><b>2-CONNECT®-LAB ANALOG</b> KS 62 = 1,6 mm</p>	<p>2CON 5.La1.600</p>	



## TITANIUM BASE

incl. screw

THO 80600000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

THO 9.S3D6.000



## NT-MILL LAB ANALOG

THO 5.La6.000



## IMPLANT PICK-UP

for open-tray with integrated screw

THO TR-00056.0



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

THO 604060



## DIM ANALOG®

THO 5.DIM.600



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



<p><b>NT-PREFORM®</b> incl. coloured milling screw Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 10.Mi6.000.05</p>	
<p><b>SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning</p>	<p>THO 9.S3D6.000</p>	
<p><b>NT-MILL LAB ANALOG</b> for NT-Preform®</p>	<p>THO 5.La6.000</p>	
<p><b>COLOURED MILLING SCREW</b> for NT-Preform® Recommended screwdriver W 11.TH0.MIL 4-socket</p>	<p>THO 6.Sc4.060</p>	
<p><b>ABUTMENT SCREW</b></p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>THO 604060</p>	
<p><b>DIM ANALOG®</b></p>	<p>THO 5.DIM.600</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>ABUTMENT GRIP</b> for NT-Holder fixture</p>	<p>W 12.300.000</p>	
<p><b>CUSTOMIZED ABUTMENT</b> milled by NT-iDent</p> <p style="color: red; font-size: small;">Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.



## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm THO-2CON 8060001S

GH 2,5 mm THO-2CON 8060002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm THO-2CON 8060001M

GH 2,5 mm THO-2CON 8060002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800





<p><b>2-CONNECT®-ABUTMENT SET (ANGLED)</b> incl. 2-CONnect®-Coping, mounting aid, Coping-screw and Male Part Screw, angled 20°, Gingiva Height 2,5 mm</p>	<p>THO-2CON 80600A3S</p>	
<p><b>2-CONNECT®-ABUTMENT</b> angled 20°, Gingiva Height 2,5 mm Male Part inclusive Screw, mounting aid</p>	<p>THO-2CON 80600A3M</p>	
<p><b>2-CONNECT®-SCAN BODY 3D-GUIDE®</b> for intra + extra oral scanning KS 62 = 1,6 mm</p>	<p>2CON 9.S3D1.600</p>	
<p><b>2-CONNECT®-COPING</b> Female Part</p>	<p>2CON F-U</p>	
<p><b>COPING-SCREW</b> Recommended tightening torques: 15 Ncm</p>	<p>KS 62</p>	
<p><b>MALE PART SCREW</b> for angled 20° 2-CONnect® Abutment Recommended tightening torques: 20 Ncm</p>	<p>THO 6.Ms4.060</p>	
<p><b>2-CONNECT®-DIM ANALOG®</b> KS 62 = 1,6 mm</p>	<p>2CON 5.DIM.160</p>	
<p><b>SLEEVE</b> (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A30</p>	
<p><b>2-CONNECT®-IMPLANT PICK-UP</b> integrated screw, KS 62 = 1,6 mm</p>	<p>2CON TR-00031.6</p>	
<p><b>2-CONNECT®-LAB ANALOG</b> KS 62 = 1,6 mm</p>	<p>2CON 5.La1.600</p>	